Certification

- Certified to EN149 + A1 2009 and FFP1.
- FFP1 Respirator masks (which means they are made of a fabric designed to filter the air or impurities) are based on a grading system do indicate how much protection they offer. First up is <u>FFP1</u>, which protects against materials in concentrations up to 4x OEL or 4x APF (assigned protection factor).
- BFE rating of 95%: The Bacterial Filtration Efficiency (**BFE**) test is performed on filtration materials and devices that are designed to provide protection against biological aerosols, such as face masks, surgical gowns, caps, and air filters.

Form QAT_10-M04, version 00, effective since March 6th, 2020

Certificate of Compliance

No. 0P200317.SDGUC48

Technical Construction File no. TPPEMJ2003123551-AG02-47

Certificate's Holder:

Certification ECM Mark:



Product: Daily Disposable Protective Mask Model(s): DF/DC 001-2020

Verification to:

Standard:

EN 149:2001+A1:2009

related to CE Directive(s):

R 2016/425 (Personal Protective Equipment)

Remark: This document has been issued on a voluntary basis and upon request of the manufacturer. It is our opinion that the technical documentation received from the manufacturer is satisfactory for the requirements of the ECM Certification Mark. The conformity mark above can be affixed on the products accordingly to the ECM regulation about its release and its use.

Additional information and clarification about the Marking:



The manufacturer is responsible for the CE Marking process. This document has been issued on the basis of the regulation on ECM Voluntary Mark for the certification of products. RG01 ECM rev.3 available at: www.entecerma.it

Issuance date: 17 March 2020

Expiry date: 16 March 2025

Reviewer Technical expert Amanda Payne



Certification

Copy of EN149:2001 + A1 2009 certificate – available on request.